



Auto Ref/Kerato/Tono/Pachymeter  
**TONOREF™ III**



**THE ART OF EYE CARE**



TONOREF III  
AUTO REF / KERATO / TONO / PACHY-METER

A MASTERPIECE  
of  
COMBINATION



The space saving TONOREF™ III  
is a comfortable and  
efficient upgrade to your practice



Auto Refractometer  
Auto Keratometer  
Non Contact Tonometer  
and  
Non Contact Pachymeter

# High Measurement Accuracy

## Refraction


### Large Pupil Zone Imaging Method

The use of a wide area measurement within the pupil increases the accuracy of measurement that is more indicative of the subjective refraction.

The large pupil zone imaging method measures the central refraction and a large area refraction.


The difference of the measurement allows assessment of the effect of pupil size such as a vision in dim light.

Measurements can be performed on small pupils as small as 2 mm.



#### **Low Confidence Alert**

The measurement ring image can be displayed to alert the operator if low confidence measurement occurs.



Ring image

### Super Luminescent Diode (SLD) Light and Highly Sensitive CCD Camera

The system combining the SLD light and highly sensitive CCD camera significantly improves measurement capability even in dense cataractous eyes.

### Optimal Fogging to Minimize Accommodation

Fogging is performed after correcting the patient's astigmatism with built-in cylinder lenses.

This minimizes the effect of accommodation even of patients with high astigmatism.

## Keratometry

### Double Mire Ring Method


Keratometry measurements performed with the mire ring method reduces interference from the eyelids.

The TONOREFT™ III performs measurements at diameters of 3.3 mm and 2.4 mm.

Comparison of the two values allows a better understanding of the cornea shape.




Measurement with double mire ring




## Pachymetry

Non-contact optical pachymetry is used to measure corneal thickness.



The principle of specular reflection for pachymetry allows a more compact design of TONOREF™ III.



The pachymetry data can be used to display a corrected IOP value.

## Tonometry

### Automated Calculation of Corrected IOP

The TONOREF™ III provides the automated calculation function of the corrected IOP based on the central corneal thickness.

Generally, the IOP is overestimated for thick corneas and underestimated for thin corneas. The corrected IOP value allows a more accurate assessment.



### Patient-friendly Air Puff

#### Automatic Puff Control (APC)

In subsequent measurements, the APC function performs the measurement with the minimum air pressure based on the previous measurement data.


#### Softer and Quieter Air Puff

The mechanical design of the TONOREF™ III reduces noise and air intensity to achieve a more gentle air puff.

#### Gentle Nozzle Design

A gentle nozzle design reduces patient's perception of physical pressure.



#### With APC



# Clinically Important Functions

## Accommodation Measurement

The accommodation measurement helps to assess such as pseudomyopia, eyestrain, and accommodative palsy. Objective measurement of accommodation is performed with patient's focusing on a target which moves from distant to near. Intelligence algorithm detects the patient's response and reduces the measurement time in patients with a slow or weak accommodative response.



## Opacity Assessment

### Retroillumination Image and NIDEK Cataract Indices

The retroillumination image allows evaluation of media opacity. NIDEK cataract indices indicate the severity of the opacity and helps to assess the progression of pathology.



Eye with light opacity



Eye with dense opacity

### NIDEK Cataract Indices

- [COI. H] Opacity size within a diameter of 3 mm in the center (vertical diameter)
- [COI. A] Opacity proportion within a diameter of 3 mm in the center
- [POI] Opacity proportion within the entire periphery

The NIDEK cataract indices are for reference only. The following conditions may indicate different indices from ones of actual status.

- Peripheral image is darkly captured due to alignment position.
- Opacities are not in focus.
- Bright spot reflecting observation light occurs on the cornea apex.
- Position of the 3 mm diameter circle is shifted due to incorrect pupil detection caused by opacity location.

# User-friendly Design

## Easy to Use Screen

- Tilttable 7-inch color LCD touchscreen
- Summary Display  
Summary screen allows easy and quick confirmation of patient data.

ALL	AR	KM	NTP	CS	PS	ACC	RETRO
-R->							
[AR]							
	S	C	A	S	C	A	
	-5.75	-0.75	172	-5.00	-1.00	9	
[KM]							
	R1	R2	deg	R1	R2	deg	
	8.22	7.92	172	8.22	7.95	11	
[NTP]							
	IOP (mmHg)	CCT (um)		IOP (mmHg)	CCT (um)		
	9.7	525		10.0	536		
	Corrected IOP (mmHg)						
	11.0			10.8			




## Easy Access to Patients Eyelids

The radical cut design allows direct access to patient eyelids.



## 3-D Auto Tracking and Auto Shot

The 3-D auto tracking and auto shot provide faster, simpler, and more accurate measurements. Once alignment is completed, the measurement starts automatically.



## Joystick for Flexible Alignment

The joystick helps the operator make fine adjustments during alignment to improve the precision, even for eyes with poor fixation which cannot be tracked with automated tracking systems.



## Quick and Easy Wireless Data Transfer

NIDEK refraction products allow for quick and easy wireless data transfer\* using the Eye Care card, WLAN or infrared communication.

This is helpful for making a simple refraction system without complicated wired connection.

\*The specifications for wireless data transfer differ according to each product and from country to country. The requirements also differ depending on the method of wireless data transfer.



## Automatic Anti Dew Heater



Automatic anti dew heater for measuring windows prevents condensation to provide accurate measurements in cooler rooms.

# TONOREF™ III Specifications

Auto refractometer	
Measurement range	Sphere -30.00 to +25.00 D (VD = 12 mm) (0.01/0.12/0.25 D increments) Cylinder 0 to ±12.00 D (0.01/0.12/0.25 D increments) Axis 0 to 180° (1°/5° increments)
Minimum measurable pupil diameter	ø2 mm
Measurement area	ø1 to 6 mm
Chart	Scenery chart
Auto keratometer	
Measurement range	Curvature radius 5.00 to 13.00 mm (0.01 mm increments) Refractive power 25.96 to 67.50 D (n = 1.3375) (0.01/0.12/0.25 D increments) Cylindrical power 0 to ±12.00 D (0.0 / 0.12/0.25 D increments) Axis 0 to 180° (1°/5° increments)
Measurement area	ø3.3 mm (R=7.7 mm), ø2.4 mm (R=7.8 mm)
Non contact tonometer	
Measurement range	1 to 60 mmHg (1 mmHg increments)
Measurement range setting	APC40, APC60 (APC=Automatic Puff Control), 40, 60
Working distance	11 mm
Eye fixation	Inner fixation light
Non contact pachymeter	
Measurement range	300 to 800 µm (1 µm increments)
IOP correction by corneal thickness	Automatic calculation
Retroillumination image	Available
Accommodation measurement range	0 to 10.00 D (0.01/0.12/0.25 D increments)
PD measurement range	30 to 85 mm (1 mm increments) (Near point PD: 28 to 80 mm at WD = 40 cm)
Corneal size measurement range	10.0 to 14.0 mm (0.1 mm increments)
Pupil size measurement range	1.0 to 10.0 mm (0.1 mm increments)
Auto tracking	X-Y-Z directions
Auto shot	Available
Display	Tilttable 7.0-inch color LCD with touch panel
Printer	Thermal line printer with easy loading and auto cutter
Interface	RS-232C: 2 port, LAN: 1 port, USB: 1 port Wireless LAN (WLAN)*: 1ch
Power supply	AC 100 to 240 V, 50/60 Hz
Power consumption	100 VA
Dimensions/Mass	260 (W) × 495 (D) × 505 (H) mm / 22 kg at ARK standard mode 260 (W) × 495 (D) × 460 (H) mm / 22 kg at NT standard mode 10.2 (W) × 19.5 (D) × 19.9 (H)" / 48 lbs. at ARK standard mode 10.2 (W) × 19.5 (D) × 18.1 (H)" / 48 lbs. at NT standard mode

\* Only for the countries (regions) certified by the Radio Law

Product/Model name: AUTO REF/KERATO/TONO/PACHYMETER TONOREF III  
Brochure and listed features of the device are intended for non-US practitioners.  
Specifications may vary depending on circumstances in each country.  
Specifications and design are subject to change without notice.



**HEAD OFFICE**  
**(International Div.)**  
34-14 Maehama,  
Hiroishi-cho, Gamagori,  
Aichi 443-0038, JAPAN  
TEL: +81-533-67-8895  
URL: www.nidek.com  
**[Manufacturer]**

**TOKYO OFFICE**  
**(International Div.)**  
3F Sumitomo Fudosan Hongo  
Bldg., 3-22-5 Hongo, Bunkyo-ku,  
Tokyo 113-0033, JAPAN  
TEL: +81-3-5844-2641  
URL: www.nidek.com

**NIDEK INC.**  
2040 Corporate Court,  
San Jose, CA 95131, U.S.A.  
TEL: +1-408-468-6400  
+1-800-223-9044  
(US Only)  
URL: usa.nidek.com

**NIDEK S.A.**  
Europarc,  
13 rue Auguste Perret,  
94042 Créteil, FRANCE  
TEL: +33-1-49 80 97 97  
URL: www.nidek.fr

**NIDEK TECHNOLOGIES S.R.L.**  
Via dell'Artigianato,  
6/A, 35020 Albignasego (Padova),  
ITALY  
TEL: +39 049 8629200/8626399  
URL: www.nidektechnologies.it

**NIDEK (SHANGHAI) CO., LTD.**  
Rm3205, Shanghai Multi  
Media Park, No.1027 Chang  
Ning Rd, Chang Ning District,  
Shanghai, CHINA 200050  
TEL: +86 021-5212-7942  
URL: www.nidek-china.cn

**NIDEK SINGAPORE PTE. LTD.**  
51 Changi Business Park  
Central 2, #06-14,  
The Signature 486066,  
SINGAPORE  
TEL: +65 6588 0389  
URL: www.nidek.sg